

# OG-6 REAGENT, PAP 2A

IVD In vitro diagnostic medical device



## Cytoplasmic staining reagent acc. to Papanicolaou Counterstain for monochromatic staining of samples in cytology

### INSTRUCTIONS FOR USE

REF Product code: OG6-OT-100 (100 mL) OG6-OT-110 (10x100 mL) OG6-OT-500 (500 mL) OG6-OT-1L (1000 mL) OG6-OT-2.5L (2500 mL)

#### Introduction

OG-6 reagent, Pap 2A is an alcohol solution of Orange G dye with added phosphomolybdic acid (PMA). The first step in using the Papanicolaou staining method implies nuclear staining with a hematoxylin solution, and next two steps consist of contrast staining using the monochromatic OG-6 reagent and one of the polychromatic EA reagent formulations consisting of two acid dyes, the Eosin Y and Light Green SF. The Orange G molecule stains the cytoplasm, and in later stages of the procedure it remains only in the mature, keratinized cells that turn different shades of orange. The third step consists of using of one of the polychromatic EA solutions that stains the unstained cellular components, such as squamous cells, nucleoli, cilia, and erythrocytes. Test samples can be gynecological and non-gynecological, such as sputum, urine, and cytological puncture samples. In order to obtain optimal staining results, OG-6 reagent, Pap 2A has properties completely in compliance with other BioGnost's reagents for cytological smearing acc. to Papanicolaou - Hematoxylin HP, Pap 1A, EA 31 reagent, Pap 3A, as well as alternative contrast polychromatic stains, such as EA 50 reagent, Pap 3B, and EA 65 reagent, Pap 3C.

#### Product description

**OG-6 REAGENT, PAP 2A** - Counterstain for monochromatic staining of samples in exfoliative cytology. Contains BSC-certified Orange G dye with added phosphomolybdic acid and necessary stabilizers.

#### Preparing the cytological smear for staining

There are two methods of collecting and preparing the cytological samples:

1. After collecting the cytological sample, place it on the microscope slide (VITROGNOST), fixate it immediately with a fixative in a spray bottle (CitoSpray), dry it and keep until the staining process. Cytological sample may be fixated and kept until staining by immersing into 95% alcohol solution (Histanol 95) for a minimum of 30 minutes.
2. Using liquid-based cytology method (LBC) and brush for collecting cytological samples, fixate the sample immediately (CitoFix, CitoFix in transport containers) by removing the brush head and immersing it in the fixative. At the beginning of processing the sample, isolate the cells from the fixative (one of the methods is to centrifuge the fixative) and place them on the microscope slide equally in a single layer. Cytological sample prepared in such a way is ready for staining.

#### The Papanicolaou staining method, PROGRESSIVE

The first stage of staining procedure depends on the method the cytological sample was collected and fixated on the microscope slide.

If the sample is dry and previously fixated by using CitoSpray, it is necessary to keep it 10% in a 95% alcohol solution (Histanol 95) for 10 minutes in order to remove polyglycols. If the section was fixated with a 95% alcohol solution (Histanol 95), ignore this step. During staining cytological samples (prepared by using the liquid based cytology method (LBC)) that contain low concentration of alcohol, rehydration by descending series of alcohol solutions is not necessary. The procedure begins by rinsing the section using distilled/demineralized water and is then stained using Hematoxylin HP.

1	Rehydration in descending series of alcohols (Histanol 95, Histanol 80 and Histanol 70) and in distilled/demineralized water	6-8 dips in each of the 4 exchanges
2	Staining using Hematoxylin HP, Pap 1A	2-3 min
3	Bluing using Scott's solution or Bluing reagent	1 min
	Note: If the mentioned reagents are not available, the section should be blued using indirect stream of water	3-5 min
4	Dehydration in ascending series of alcohols (Histanol 70, Histanol 80 and Histanol 95)	6-8 dips in each of the 3 exchanges
5	Staining using OG-6 reagent, Pap 2A	2-3 min
6	Rinsing using 95% alcohol in two exchanges (Histanol 95)	6-8 dips in each of the 2 exchanges
7	Staining using EA 31 reagent, Pap 3A or EA 50 reagent, Pap 3B	2-3 min
8	Rinsing using 95% alcohol (Histanol 95)	6-8 dips
9	Dehydration using 100% alcohol (Histanol 100)	6-8 dips
10	Dehydration using 100% alcohol (Histanol 100)	3-5 min
11	Clear the section in xylene (BioClear) or in a xylene substitute (BioClear New)	6-8 dips
12	Clear the section in xylene (BioClear) or in a xylene substitute (BioClear New)	3-5 min

Immediately after clearing apply an appropriate BioMount medium for covering/mounting on the section. If BioClear xylene was used, use one of BioGnost's mounting xylene-based media (BioMount, BioMount High, BioMount M, BioMount DPX, BioMount C, or universal BioMount New). If BioClear New xylene substitute was used, the appropriate covering agent is BioMount New. Cover the section with a VitroGnost cover glass.

#### Note

In the case of subsidence in the Hematoxylin HP, Pap 1A or formation of metallic glow on the surface, reagent should be filtrated before use Time periods of staining procedures are not completely standardized. The suggested methods are in accordance with BioGnost's reagents' properties and correspond to longtime clinical and laboratory practice. Intensity of staining depends on the period of exposure to stains and reagents. Staining procedure can be changed according to personal preferences if they correspond to the basic principles of cytotechnology.

## Papanicolaou staining method, **REGRESSIVE**

The regressive staining method creates a better sample differentiation and clearer nuclear structure visibility.

The first stage of staining procedure depends on the method the cytological sample was collected and fixated on the microscope slide.

If the sample is dry and previously fixated by using CitoSpray, it is necessary to keep it 10% in a 95% alcohol solution (Histanol 95) for 10 minutes in order to remove polyglycols. If the section was fixated with a 95% alcohol solution (Histanol 95), ignore this step. During staining cytological samples (prepared by using the liquid based cytology method (LBC)) that contain low concentration of alcohol, rehydration by descending series of alcohol solutions is not necessary. The procedure begins by rinsing the section using distilled/demineralized water and is then stained using Hematoxylin HP.

1	Rehydration in descending series of alcohols (Histanol 95, Histanol 80 and Histanol 70) and in distilled/demineralized water	6-8 dips in each of the 4 exchanges
2	Staining using Hematoxylin HP, Pap 1A	6 min
3	Rinsing in distilled/demineralized water	6-8 dips
4	Differentiation using HCL Pap reagent or in 0.1% HCl solution	5-10 seconds
	Note: This step removes excessive hematoxylin from the nucleus and cytoplasm. Discoloration of the nuclei can occur if the section is treated with the differentiation agent for too long	
5	Rinsing in distilled water	6-8 dips
6	Bluing using Scott's solution or Bluing reagent	1 min
	Note: If the mentioned reagents are not available, the section should be blued using indirect stream of water	3-5 min
7	Dehydration in ascending series of alcohols (Histanol 70, Histanol 80 and Histanol 95)	6-8 dips in each of the 3 exchanges
8	Staining using OG-6 reagent, Pap 2A	3 min
9	Rinsing using 95% alcohol in two exchanges (Histanol 95)	6-8 dips in each of the 2 exchanges
10	Staining using EA 31 reagent, Pap 3A or EA 50 reagent, Pap 3B	3 min
11	Rinsing using 95% alcohol (Histanol 95)	6-8 dips
12	Dehydration using 100% alcohol (Histanol 100)	6-8 dips
13	Dehydration using 100% alcohol (Histanol 100)	3-5 min
14	Clear the section in xylene (BioClear) or in a xylene substitute (BioClear New)	6-8 dips
15	Clear the section in xylene (BioClear) or in a xylene substitute (BioClear New)	3-5 min

Immediately after clearing apply an appropriate BioMount medium for covering/mounting on the section. If BioClear xylene was used, use one of BioGnost's mounting xylene-based media (BioMount, BioMount High, BioMount M, BioMount DPX, BioMount C, or universal BioMount New). If BioClear New xylene substitute was used, the appropriate covering agent is BioMount New. Cover the section with a VitroGnost cover glass.

### Note

In the case of subsidence in the Hematoxylin HP, Pap 1A or formation of metallic glow on the surface, reagent should be filtrated before use Time periods of staining procedures are not completely standardized. The suggested methods are in accordance with BioGnost's reagents' properties and correspond to longtime clinical and laboratory practice. Intensity of staining depends on the period of exposure to stains and reagents. Staining procedure can be changed according to personal preferences if they correspond to the basic principles of cytotechnology.

### Result

Blue - nuclei

Yellow-orange - keratinized cells

Pink-red - superficial squamous epithelial cell, erythrocytes, nucleoli, cilia

Green - cytoplasm of all the other cell types (parabasal and intermediate squamous cells, columnar cells, polymorphonuclear leukocytes, lymphocytes, histiocytes, adenocarcinomas, undifferentiated carcinoma cells)

### Preparing the sample and diagnostics

Use only appropriate instruments for collecting and preparing the samples. Process the samples with modern technology and mark them clearly. Follow the manufacturer's instructions for handling. In order to avoid mistakes, the staining procedure and diagnostics should only be conducted by authorized and qualified personnel. Use only microscope according to standards of the medical diagnostic laboratory.

### Safety at work and environmental protection

Handle the product in accordance with safety at work and environmental protection guidelines. Used solutions and out of date solutions should be disposed of as special waste in accordance with national guidelines. Chemicals used in this procedure could pose danger to human health. Tested tissue specimens are potentially infectious. Necessary safety measures for protecting human health should be taken in accordance with good laboratory practice. Act in accordance with signs and warnings notices printed on the product's label, as well as in BioGnost's material safety data sheet.














### Storing, stability and expiry date

Keep OG-6 reagent, Pap 2A in a tightly closed original package at temperature between 15°C and 25°C. Keep in dry places, do not freeze and avoid exposing to direct sunlight. Date of manufacture and expiry date are printed on the product's label.

### References

1. Papanicolaou, G.N. (1941): Some improved methods for staining vaginal smears. J Lab Clin Med.
2. Papanicolaou, G.N. (1942): A new procedure for staining vaginal smears. Science.
3. Carson, F.L., Hladik C. (2009): Histotechnology: A self-instructional text, 3<sup>rd</sup> ed. ASCP Press.
4. Sherwani, R.K., Khaqn, T. et al. (2007): Conventional Pap Smear and Liquid Based Cytology for Cervical Cancer Screening – A Comparative Study, Journal of Cytology, 24 (4): pp 167-172.

OG6-OT-X, V25-EN9, 03 November 2015, IŠP/VR

	Refer to the supplied documentation		Storage temperature range		Number of tests in package		Product code		European Conformity
	Refer to supplied instructions		Keep away from heat and sunlight		Valid until		Lot number		Manufacturer
	For <i>in vitro</i> diagnostic use only		Keep in dry place		Caution - fragile				

 BIOGNOST Ltd.  
Medjugorska 59  
10040 Zagreb  
CROATIA  
www.biognost.com

